

MEYRA GmbH
Meyra-Ring 2
32689 Kalletal
Deutschland

Tel.: +49 57 33 922- 0
Fax: +49 57 33 922-143
info@meyra.de
www.meyra.de

MEYRA GmbH • Meyra-Ring 2 • D-32689 Kalletal

Adresse

Ihr Zeichen, Ihre Nachricht vom

Unser Zeichen
MN

Durchwahl
120

Ihr Gesprächspartner
Marcus Nolting

Datum
17.06.2021

Urgent safety information

PM 2021/1

Preventive replacement of the seat carrier for the electric vehicles

**CL409 (model no. 1.254) and
CL510 (model no.1.264)**

from the delivery period: 6/8/2020 - 4/30/2021

Dear Sirs and Madames,

As part of our market observation, we became aware of a possible weak point in our electric vehicles CL409 and CL 510. The seat carrier on these vehicles could break suddenly and as a result the user could be seriously injured. In this context, it is recommended, that the vehicles will no longer be used, until the seat carrier has been replaced.

With this cover letter you will receive a list of the vehicles affected, as well as a request form with which you can order the required seat carriers from us free of charge after checking your data records.

Actions you should take:

- Inform your customers about the corrective action
- Feedback to Meyra, that your customers have been informed
- Requesting the required seat carriers from customer service
- Replacement of the affected seat carriers at your customer vehicles

Hausanschrift:
Meyra-Ring 2
32689 Kalletal-Kalldorf
Deutschland

MEYRA GmbH
Sitz: Kalletal-Kalldorf
AG Lemgo HRB 8117

Bankverbindung:
Commerzbank Herford
IBAN: DE22 4944 0043 0290 0058 00
BIC: COBADEFF494

Geschäftszeiten:
Mo. – Do. 08.00 – 17.00 Uhr
Fr. 08.00 – 15.00 Uhr

Steuer-Nr.: 329/5744/2042
Ust-IdNr.: DE815455949

Geschäftsführer:
Tadeusz Baczyński, Frank B. McDowell, Friedrich Kottmeier, Stefan Schäfer

Unser Unternehmen ist zertifiziert nach DIN EN ISO
13485:2016 WEEE-Reg.-Nr. DE 73325670

Passing on the information described:

Please ensure, that all users of the affected products and other persons are informed about this urgent safety information. If you have given the products to third parties, please forward a copy of this information and inform our contact person given below.

We would like to thank you in advance for your efforts.

If you have any further questions, Mr. Marcus Nolting will be happy to answer by phone +49 (0) 57333 / 922-120 or by email at Marcus.Nolting@Meyra.de.

All of this information and attachments are also available in electronic form on our website at www.meyra.com/sicherheitshinweise-rueckrufe/

With best regards

MEYRA GmbH



Managing director



MDR-Responsible person

Appendix 1: List of affected products

Appendix 2: Request form

Appendix 3: Confirmation form

Appendix 4: Assembly instructions